



JUP-1002-001

The Break-Away is designed with the busy medical environment in mind. With the new Break-Away in use, the patient station is relieved of stress if pressure is inadvertently applied. The design will save on costly repairs to the patient station and pillow speaker. When pressure is applied, the Break-Away will separate from the pillow speaker and place a call on the master station to notify the caregiver that the pillow speaker has become separated. The new West-Call® Break-Away cord is designed with a medical-style connector that adapts to the West-Call® pillow speaker jack.

- ▶ Rugged style
- ▶ Medical-style 12-pin connector to connect to the mating pillow speaker medical-style connector
- ▶ 10-pin RJ style to connect to the West-Call® Patient Station
- ▶ Keyed connector to prevent improper connection

VERSIONS AND MODEL NUMBERS:

STATION TYPE	MODEL NO.
Break-Away for Pillow Speaker has a 1010 male to 1012 female connector. To be used with pillow speakers with a 1012 connector	JUP-1002-001
To be used in conjunction with the Break-Away, where the pillow speaker has the tab on the 10-pin connector	JUP-1001-001

- For use with the West-Call® Novus®, Odyssey and Endeavor systems

- ▶ 10-pin to medical connector
- ▶ Rated at seven pounds of pull
- ▶ Termination: Keyed connectors

Certifications UL-1069 for US and Canada, E139082

NOTE: This information should not be used for final design. Consult West-Com for total system requirements. Specifications subject to change without notice.



View West-Com's complete West-Call® line of healthcare communications solutions at www.westcall.com

West-Com Nurse Call Systems, Inc.
 2200 Cordelia Road • Fairfield, CA 94534 • 800-761-1180 • info@westcall.com