



EV-ASM/EXT-1

The External Audio Interface Board for the West-Call® Endeavor nurse call system provides 8 external audio channels for use with wireless telephone interface both Novus® and Endeavor audio systems. Must be used in conjunction with an EV/MMC 4x8 or 8x16.

Features:

- ▶ 8 External Audio Channels
- ▶ Integration to patient station using Wireless Telephone Interface

System Requirements:

- ▶ Endeavor Multi-Master Controller (EV-MMC)
- ▶ 1822/6 Power Supply Can for mounting EV-MMC
- ▶ WTI-4 or WTI-8 for the Wireless Telephone Interface

VERSIONS AND MODEL NUMBERS:

STATION TYPE	MODEL NO.
Endeavor Audio Module for up to 4 masters and 8 audio paths. Includes 1 CNTR1, 1 EV-ASM/Base, 1 EV-ASM/EXT1, and 1 Power Supply. Enclosure can and door must be ordered separately.	EV-MMC-4x8/EXT1
Endeavor Audio Module for up to 8 masters and 16 audio paths. Includes 1 CNTR1, 1 EV-ASM/Base, 1 EV-ASM/EXP1, 1 EV-ASM/EXT1, and 1 Power Supply. Enclosure can and door must be ordered separately.	EV-MMC-8x16/EXT1
External Audio Interface Board	EV-ASM/EXT1
Surface Mount Cabinet with Door	TC-1822/6 and TCD-1822/S
Flush Mount Cabinet with Door	TC-1822/6 and TCD-1822/F

- For use with the West-Call® Endeavor system

Operating Voltage	5VDC, powered from EV-MMC
Size	8 ½" x 3 ½"
Termination	8 - 3 Pin 100 Spacing
Cable Requirements	22/4 Cond. to WTI-4 or WTI-8 per External Channel
Certifications	UL-1069 for US and Canada, E139082

NOTE: This information should not be used for final design. Consult West-Com for total system requirements. Specifications subject to change without notice.



View West-Com's complete West-Call® line of healthcare communications solutions at www.westcall.com

West-Com Nurse Call Systems, Inc.
2200 Cordelia Road • Fairfield, CA 94534 • 800-761-1180 • info@westcall.com