



NOVUS® VISUAL PATIENT STATION

The single-gang Visual Patient Station is designed to be used with the West-Call® Novus® and Novus® Visual systems. When used with the West-Call® Novus® Visual System, it is a stand-alone patient station. The VPS includes the standard ¼ " phono receptacle for use with a call cord, CC96-030 or CC144-100. The patient uses the call cord to signal for assistance. A call is automatically placed when the call cord is removed. When used with the West-Call® Novus®, the NV VPS must be used in conjunction with the SDL2, NV-Dome or Novus® LDL2. When used with the West-Call® Novus® Visual, the NVV-VPS must be used in conjunction with a NV-VDL or NV-VDL/4.

- ▶ Standard ¼ " phono socket for a call cord used by the patient to signal for assistance
- ▶ Automatic placement of call when call cord is removed
- ▶ Highly visible call LED to assure that a call has been placed
- ▶ Large "CANCEL" button for cancellation of calls
- ▶ Fire retardant high-impact hospital grade molded thermoplastic with inlaid membrane
- ▶ Attractive Lexan membrane (Pantone 1C Warm Grey)
- ▶ Solid state circuitry
- ▶ Easy plug-in connection into the system wiring

VERSIONS AND MODEL NUMBERS:

STATION TYPE	NOVUS®	NOVUS® VISUAL
Novus® Visual Patient Station	NV-VPS	NVV-VPS
AUXILIARY EQUIPMENT		
Call Cord Assembly (8 foot)	CC96-030	CC96-030
Call Cord Assembly (12 foot)	CC144-100	CC144-100

- For use with the West-Call® Novus® system
- Custom color faceplate available upon request
- Custom color membrane available upon request
- Custom membrane wording available upon request

Operating Voltage	28 VDC
Power Consumption	1 watt maximum
Dimensions	4 ½" H x 2 ¾" W x 1 ½" D
Mounting	4-inch square backbox with one-gang plaster ring, Raco 190 with 768 one-gang adapter or equivalent. Minimum 2 1/8" depth required. Normal mounting is vertical. For remodel areas, use Raco #500 or 503 with #977 Grip-Lok's or equivalent. Minimum 2 ½" depth required
Faceplate	Hospital grade UL94V-0 ABS molded thermoplastic with inlaid membrane
Faceplate Color	Standard color is Pantone 1C Warm Grey
Membrane	.010 Lexan with hardcoat finish background color Pantone 4C Warm Grey Cancel color Pantone 1C and 11C Warm Grey
Lamp Life	500,000 hour minimum
Switch Life	100,000 cycles minimum
Call Cord Socket	1/4" Phono
Cable Requirement	For NV-VPS, use CAT5 cable to associated dome light. For NVV-VPS, use Cat-5 Cable from the previous device and 4 conductor #22AWG cable to associated NV-VDL
Termination Requirements	RJ-45 connector used on NV-VPS. NVV-VPS uses Panduit CE100F22-4 and RJ-45
Certifications	UL-1069 for US and Canada, E139082

NOTE: This information should not be used for final design. Consult West-Com for total system requirements. Specifications subject to change without notice.



View West-Com's complete West-Call® line of healthcare communications solutions at www.westcall.com

West-Com Nurse Call Systems, Inc.
2200 Cordelia Road • Fairfield, CA 94534 • 800-761-1180 • info@westcall.com