



**VISUAL PATIENT
STATION**

The single gang Visual Patient Station is designed to be used with the West-Call® Odyssey Plus and West-Call® Endeavor systems. When used with the West-Call® Endeavor system, it is a stand-alone patient station. The VPS includes the standard 1/4" phono receptacle for use with a call cord, CC96-030 or CC144-100. The patient uses the call cord to signal for assistance. A call is automatically placed when the call cord is removed. When used with the West-Call® Odyssey Plus, the VPS-1000 must be used in conjunction with the SDL-1000 (Smart Dome Light) or a Closure Plus Module.

- ▶ Standard 1/4" phono socket for a call cord used by the patient to signal for assistance
- ▶ Automatic placement of call when call cord is removed
- ▶ Highly visible call LED to assure that a call has been placed
- ▶ Large color-coded "CANCEL" button for cancellation of calls
- ▶ Fire retardant high-impact hospital grade molded thermoplastic with inlaid membrane
- ▶ Attractive Lexan membrane (Pantone 1C Warm Grey)
- ▶ Solid state circuitry
- ▶ Easy plug-in connection into the system wiring

DESCRIPTION	ODYSSEY PLUS MODEL NO.	ENDEAVOR MODEL NO.
Visual Patient Station	VPS-1000	EV VPS

AUXILIARY EQUIPMENT

Call Cord Assembly (8 foot)	CC96-030	CC96-030
Call Cord Assembly (12 foot)	CC144-100	CC144-100
Smart Dome Light (for use with West-Call® Odyssey Plus)	SDL-1000	
Closure Plus Module (for use with the West-Call® Odyssey Plus)	CL+	

- Custom color faceplate available upon request
- Custom color membrane available upon request
- Custom membrane wording available upon request
- For use with the West-Call® Odyssey Plus and West-Call® Endeavor systems

Operating Voltage	12-30 VDC; 12-18 VAC
Power Consumption	1 watt maximum
Dimensions	4 1/2"H x 2 3/4"W x 1 1/2"D
Mounting	4-inch square backbox with one-gang plaster ring, Raco 190 with 768 one-gang adapter or equivalent. Minimum 2 1/8" depth required. Normal mounting is vertical. For remodel areas, use Raco #500 or 503 with #977 Grip-Lok's or equivalent. Minimum 2 1/2" depth required
Faceplate	Hospital grade UL94V-0 ABS molded thermoplastic with inlaid membrane
Faceplate Color	Standard color is Pantone 1C Warm Grey
Membrane	.010 Lexan with hardcoat finish background color Pantone 4C Warm Grey Cancel color Pantone 1C and 11C Warm Grey
Lamp Life	500,000 hour minimum
Switch Life	100,000 cycles minimum
Call Cord Socket	1/4" Phono
Cable Requirement	Use 4 conductor #22AWG cable to associated patient station or SDL-1000. Use 6 conductor #22AWG cable to associated Closure Plus Modules or I/O board
Termination Requirements	Plug on connector allows rapid mass termination. Mating connector is a Panduit CE156F22-10



View West-Com's complete West-Call® line of healthcare communications solutions at www.westcall.com

West-Com Nurse Call Systems Incorporated
P.O. Box 640 • Danville, CA 94526 • 800-761-1180 • fax: 925-831-0842 • info@westcall.com